

file name: 94100114

sized: 10/28/1994

meter identification: 94100114 tag:
 application customer: ARGONNE LABS
 V-Cone model number: V000B-AL

fluid: Water/water

comments: _____

fluid boiling point: 671.800 R

Operating Conditions:
 temperature: 75.000 F
 pressure: 150.000 psig
 fluid mass: 62.098#/ft³
 viscosity: 0.834 cP
 Z-factor: 0.003

CERTIFIED: *JAM*

Meter Specifications:
 inside diameter: 0.656in
 cone diameter: 0.468in
 beta ratio: 0.7007
 Fa: 1.0001

turndown: 10.0 to 1 Cf: 0.8060
 - Cf is linear to flowrate -

FLOWRATE (GPM)	DIFF-PRESS (inWC)	VELOCITY (f/s)	REYNOLDS NUMBER	COEFFICIENT OF FLOW
8.0000	51.819	7.59	45927.	0.80600
6.9022	38.573	6.55	39625.	0.80600
6.3925	33.087	6.07	36699.	0.80600
5.9720	28.877	5.67	34285.	0.80600
5.5929	25.328	5.31	32109.	0.80600
5.2363	22.201	4.97	30061.	0.80600
4.8916	19.374	4.64	28082.	0.80600
4.5516	16.774	4.32	26130.	0.80600
4.2105	14.354	4.00	24172.	0.80600
3.8628	12.082	3.67	22176.	0.80600
3.5024	9.9321	3.32	20107.	0.80600
3.1212	7.8876	2.96	17918.	0.80600
2.7072	5.9342	2.57	15542.	0.80600
2.2394	4.0605	2.13	12856.	0.80600
1.6698	2.2577	1.58	9586.4	0.80600
0.79997	0.51815	0.759	4592.5	0.80600

sized by *Candy Parker*, *10/28/94*, approval: _____