

file name: 94100124

sized: 10/28/1994

meter identification: 94100124 tag:
 application customer: ARGONNE LABS
 V-Cone model number: V000B-AL

fluid: Water/water

comments: _____

fluid boiling point: 671.800 R

Operating Conditions:

temperature: 75.000 F
 pressure: 150.000 psiG
 fluid mass: 62.098#/ft3
 viscosity: 0.834 cP
 Z-factor: 0.003

CERTIFIED: <u>JAM</u>
<u>10/31/94</u>

Meter Specifications:

inside diameter: 0.656in
 cone diameter: 0.468in
 beta ratio: 0.7007
 Fa: 1.0001

turndown: 10.0 to 1 Cf: 0.8110
 Cf is linear to flowrate -

FLOWRATE (GPM)	DIFF-PRESS (inWC)	VELOCITY (f/s)	REYNOLDS NUMBER	COEFFICIENT OF FLOW
7.9999	51.181			
6.9021	38.098	7.59	45926.	0.81100
6.3924	32.679	6.55	39624.	0.81100
5.9719	28.521	6.07	36698.	0.81100
5.5928	25.015	5.67	34284.	0.81100
5.2362	21.927	5.31	32108.	0.81100
4.8915	19.135	4.97	30061.	0.81100
4.5515	16.567	4.64	28082.	0.81100
4.2104	14.177	4.32	26130.	0.81100
3.8628	11.933	4.00	24172.	0.81100
3.5023	9.8097	3.67	22176.	0.81100
3.1211	7.7904	3.32	20107.	0.81100
2.7072	5.8610	2.96	17918.	0.81100
2.2394	4.0105	2.57	15542.	0.81100
1.6698	2.2299	2.13	12856.	0.81100
0.79995	0.51176	1.58	9586.2	0.81100
		0.759	4592.5	0.81100

Sized by: Cathy Jaker, 10/31/94, approval: _____