

file name: 94100341

sized: 11/28/1994

meter identification: 94100341 tag:
 application customer: ARGONNE LABS.
 V-Cone model number: V000A-*AL*

fluid: Water/water

comments: _____

fluid boiling point: 671.800 R

Operating Conditions:

temperature: 75.000 F
 pressure: 150.000 psig
 fluid mass: 62.098#/ft3
 viscosity: 0.834 cP
 Z-factor: 0.003

CERTIFIED: <i>JAM</i>

Meter Specifications:

inside diameter: 0.438in
 cone diameter: 0.371in
 beta ratio: 0.5315
 Fa: 1.0001

turndown: 10.0tol Cf: 0.8290
 - Cf is linear to flowrate -

FLOWRATE (GPM)	DIFF-PRESS (inWC)	VELOCITY (f/s)	REYNOLDS NUMBER	COEFFICIENT OF FLOW
1.9999	56.411	4.26	17195.	0.82900
1.7254	41.992	3.67	14836.	0.82900
1.5980	36.019	3.40	13740.	0.82900
1.4929	31.436	3.18	12836.	0.82900
1.3981	27.572	2.98	12021.	0.82900
1.3090	24.168	2.79	11255.	0.82900
1.2228	21.090	2.60	10514.	0.82900
1.1378	18.260	2.42	9783.2	0.82900
1.0525	15.626	2.24	9050.1	0.82900
0.96564	13.152	2.06	8302.8	0.82900
0.87553	10.812	1.86	7528.0	0.82900
0.78023	8.5866	1.66	6708.7	0.82900
0.67675	6.4600	1.44	5818.9	0.82900
0.55981	4.4204	1.19	4813.4	0.82900
0.41743	2.4577	0.889	3589.2	0.82900
0.19998	0.56406	0.426	1719.4	0.82900

sized by: *Candy Jula*, *11/29/94* approval: _____