

file name:94102242

sized:12/20/1994

meter identification: 94102242 tag:
 application customer: ARGONNE LABS
 V-Cone model number: V000A-4L


fluid: Water/water

comments: _____

fluid boiling point: 671.800 R

Operating Conditions:

temperature: 75.000 F
 pressure: 150.000 psiG
 fluid mass: 62.098#/ft3
 viscosity: 0.834 cP
 Z-factor: 0.003

CERTIFIED: _____


Meter Specifications:

inside diameter: 0.438in
 cone diameter: 0.371in
 beta ratio: 0.5315
 Fa: 1.0001

turndown: 10.0tol Cf: 0.8140
 - Cf is linear to flowrate -

FLOWRATE (GPM)	DIFF-PRESS (inWC)	VELOCITY (f/s)	REYNOLDS NUMBER	COEFFICIENT OF FLOW
1.9999	58.510	4.26	17195.	0.81400
1.7254	43.554	3.67	14836.	0.81400
1.5980	37.359	3.40	13740.	0.81400
1.4929	32.605	3.18	12836.	0.81400
1.3981	28.597	2.98	12022.	0.81400
1.3090	25.067	2.79	11255.	0.81400
1.2228	21.875	2.60	10514.	0.81400
1.1378	18.940	2.42	9783.2	0.81400
1.0525	16.207	2.24	9050.1	0.81400
0.96564	13.641	2.06	8302.8	0.81400
0.87553	11.214	1.86	7528.1	0.81400
0.78023	8.9060	1.66	6708.7	0.81400
0.67676	6.7003	1.44	5818.9	0.81400
0.55981	4.5848	1.19	4813.4	0.81400
0.41743	2.5492	0.889	3589.2	0.81400
0.19998	0.58505	0.426	1719.5	0.81400

sized by Carol Parker, 12/22/94, approval: _____