

file name:94100216

sized:11/07/1994

meter identification: 94100216 tag:  
 application customer: ARGONNE LABS  
 V-Cone model number: V000A-AL

fluid: Water/water comments:

fluid boiling point: 671.800 R

Operating Conditions:  
 temperature: 75.000 F  
 pressure: 150.000 psiG  
 fluid mass: 62.098#/ft3  
 viscosity: 0.834 cP  
 Z-factor: 0.003

CERTIFIED: JAM

Meter Specifications:  
 inside diameter: 0.438in  
 cone diameter: 0.299in  
 beta ratio: 0.7307  
 Fa: 1.0001

turndown: 10.0to1 Cf: 0.7430  
 - Cf is linear to flowrate -

FLOWRATE ( GPM )	DIFF-PRESS ( inWC )	VELOCITY ( f/s )	REYNOLDS NUMBER	COEFFICIENT OF FLOW
4.0002	61.106	8.52	34394.	0.74300
3.4512	45.486	7.35	29675.	0.74300
3.1964	39.016	6.80	27483.	0.74300
2.9861	34.052	6.36	25675.	0.74300
2.7966	29.867	5.95	24046.	0.74300
2.6183	26.179	5.57	22512.	0.74300
2.4459	22.846	5.21	21030.	0.74300
2.2759	19.780	4.85	19569.	0.74300
2.1053	16.927	4.48	18102.	0.74300
1.9315	14.247	4.11	16607.	0.74300
1.7513	11.712	3.73	15058.	0.74300
1.5606	9.3012	3.32	13419.	0.74300
1.3537	6.9977	2.88	11639.	0.74300
1.1198	4.7882	2.38	9627.9	0.74300
0.83495	2.6623	1.78	7179.1	0.74300
0.40000	0.61101	0.852	3439.3	0.74300

sized by: Cathy Parker 11/11/94 approval: \_\_\_\_\_