

file name:94102227

sized:12/20/1994

meter identification: 94102227 tag:
 application customer: ARGONNE LABS
 V-Cone model number: V000A-AL

fluid: Water/water

comments: _____

fluid boiling point: 671.800 R

Operating Conditions:

temperature: 75.000 F
 pressure: 150.000 psiG
 fluid mass: 62.098#/ft3
 viscosity: 0.834 cP
 Z-factor: 0.003

CERTIFIED:

Meter Specifications:

inside diameter: 0.438in
 cone diameter: 0.299in
 beta ratio: 0.7307
 Fa: 1.0001

turndown: 10.0to1 Cf: 0.7560
 - Cf is linear to flowrate -

FLOWRATE (GPM)	DIFF-PRESS (inWC)	VELOCITY (f/s)	REYNOLDS NUMBER	COEFFICIENT OF FLOW
4.0002	59.023	8.52	34394.	0.75600
3.4512	43.936	7.35	29675.	0.75600
3.1964	37.686	6.80	27483.	0.75600
2.9861	32.891	6.36	25675.	0.75600
2.7966	28.848	5.95	24046.	0.75600
2.6183	25.287	5.57	22512.	0.75600
2.4459	22.067	5.21	21030.	0.75600
2.2759	19.106	4.85	19569.	0.75600
2.1053	16.350	4.48	18102.	0.75600
1.9315	13.761	4.11	16607.	0.75600
1.7513	11.313	3.73	15058.	0.75600
1.5606	8.9841	3.32	13419.	0.75600
1.3537	6.7591	2.88	11639.	0.75600
1.1198	4.6250	2.38	9627.9	0.75600
0.83495	2.5715	1.78	7179.1	0.75600
0.40000	0.59018	0.852	3439.3	0.75600

sized by: Quality Factor, 12/22/94 approval: _____